

# OCULUS Smartfield



DICOM CONFORMANCE STATEMENT  
for Smartfield 3.19 (2023-06-19)

# 1 Conformance Statement Overview

The OCULUS Smartfield is used for measuring the visual field of the human eye. The OCULUS Smartfield supports several DICOM Service Classes, using the OFFIS DICOM Toolkit (DCMTK), to provide the following capabilities:

- Query an information system for worklists of procedures to be performed (including the patient demographics).
- Send DICOM objects created by the system to a remote system.

Table 1.1 provides an overview of the network services supported by OCULUS Smartfield . The system does not support any media services.

Table 1.1: Network Services

| Networking SOP Classes                                        | User of Service (SCU) | Provider of Service (SCP) |
|---------------------------------------------------------------|-----------------------|---------------------------|
| <b>Transfer</b>                                               |                       |                           |
| Encapsulated PDF Storage                                      | Yes                   | No                        |
| Secondary Capture Image Storage                               | Yes                   | No                        |
| Ophthalmic Visual Field Static Perimetry Measurements Storage | Yes                   | No                        |
| <b>Workflow Management</b>                                    |                       |                           |
| Modality Worklist Information Model - FIND                    | Yes                   | No                        |

# Contents

|           |                                                               |          |
|-----------|---------------------------------------------------------------|----------|
| <b>1</b>  | <b>Conformance Statement Overview</b>                         | <b>2</b> |
| <b>2</b>  | <b>Introduction</b>                                           | <b>5</b> |
| 2.1       | Revision History . . . . .                                    | 5        |
| 2.2       | Audience . . . . .                                            | 5        |
| 2.3       | Remarks . . . . .                                             | 5        |
| 2.4       | Terms and Definitions . . . . .                               | 6        |
| 2.5       | Basics of DICOM Communication . . . . .                       | 7        |
| 2.6       | Abbreviations . . . . .                                       | 8        |
| 2.6.1     | References . . . . .                                          | 8        |
| <b>3</b>  | <b>Networking</b>                                             | <b>9</b> |
| 3.1       | Implementation Model . . . . .                                | 9        |
| 3.1.1     | Application Data Flow . . . . .                               | 9        |
| 3.1.2     | Functional Definitions of AEs . . . . .                       | 9        |
| 3.1.2.1   | WORKLIST-SCU . . . . .                                        | 9        |
| 3.1.2.2   | STORAGE-SCU . . . . .                                         | 9        |
| 3.1.3     | Sequencing of Real-World Activities . . . . .                 | 10       |
| 3.2       | AE Specifications . . . . .                                   | 10       |
| 3.2.1     | WORKLIST-SCU . . . . .                                        | 10       |
| 3.2.1.1   | SOP Classes . . . . .                                         | 10       |
| 3.2.1.2   | Association Policies . . . . .                                | 10       |
| 3.2.1.2.1 | General . . . . .                                             | 10       |
| 3.2.1.2.2 | Number of Associations . . . . .                              | 10       |
| 3.2.1.2.3 | Asynchronous Nature . . . . .                                 | 10       |
| 3.2.1.2.4 | Implementation Identifying Information . . . . .              | 10       |
| 3.2.1.3   | Association Initiation Policy . . . . .                       | 10       |
| 3.2.1.3.1 | Activity - System sends a query and shows matches . . . . .   | 11       |
| 3.2.1.4   | Association Acceptance Policy . . . . .                       | 13       |
| 3.2.2     | STORAGE-SCU . . . . .                                         | 14       |
| 3.2.2.1   | SOP Classes . . . . .                                         | 14       |
| 3.2.2.2   | Association Policies . . . . .                                | 14       |
| 3.2.2.2.1 | General . . . . .                                             | 14       |
| 3.2.2.2.2 | Number of Associations . . . . .                              | 14       |
| 3.2.2.2.3 | Asynchronous Nature . . . . .                                 | 14       |
| 3.2.2.2.4 | Implementation Identifying Information . . . . .              | 14       |
| 3.2.2.3   | Association Initiation Policy . . . . .                       | 14       |
| 3.2.2.3.1 | Activity - System creates new instance and sends it . . . . . | 14       |
| 3.2.2.4   | Association Acceptance Policy . . . . .                       | 17       |
| 3.3       | Physical Network Interfaces . . . . .                         | 17       |
| 3.3.1     | Physical Network Interface . . . . .                          | 17       |
| 3.3.2     | Additional Protocols . . . . .                                | 17       |
| 3.3.3     | IPv4 and IPv6 Support . . . . .                               | 17       |
| 3.4       | Configuration . . . . .                                       | 17       |
| 3.4.1     | AE Title / Presentation Address Mapping . . . . .             | 17       |
| 3.4.1.1   | Local AE Titles . . . . .                                     | 18       |

|                                                                                 |           |
|---------------------------------------------------------------------------------|-----------|
| 3.4.1.2 Remote AE Titles . . . . .                                              | 18        |
| 3.4.2 Configurable Parameters . . . . .                                         | 18        |
| <b>4 Media Storage</b>                                                          | <b>19</b> |
| <b>5 Support of extended character sets</b>                                     | <b>20</b> |
| <b>6 Security</b>                                                               | <b>21</b> |
| 6.1 Security Profiles . . . . .                                                 | 21        |
| 6.2 Association Level Security . . . . .                                        | 21        |
| 6.3 Application Level Security . . . . .                                        | 21        |
| <b>7 Annexes</b>                                                                | <b>22</b> |
| 7.1 IOD Contents . . . . .                                                      | 22        |
| 7.1.1 Created SOP Instances . . . . .                                           | 22        |
| 7.1.1.1 Encapsulated PDF IOD . . . . .                                          | 22        |
| 7.1.1.2 Secondary Capture Image IOD . . . . .                                   | 23        |
| 7.1.1.3 Ophthalmic Visual Field Static Perimetry Measurements IOD . . . . .     | 24        |
| 7.1.1.4 Common Modules . . . . .                                                | 30        |
| 7.1.1.5 Encapsulated PDF Modules . . . . .                                      | 32        |
| 7.1.1.6 Secondary Capture Image Modules . . . . .                               | 34        |
| 7.1.1.7 Ophthalmic Visual Field Static Perimetry Measurements Modules . . . . . | 36        |
| 7.1.2 Usage of Attributes from Received IODs . . . . .                          | 40        |
| 7.1.3 Attribute Mapping . . . . .                                               | 40        |
| 7.1.4 Coerced / Modified Fields . . . . .                                       | 40        |
| 7.2 Data Dictionary of Private Attributes . . . . .                             | 40        |
| 7.3 Coded Terminology and Templates . . . . .                                   | 40        |
| 7.4 Grayscale Image Consistency . . . . .                                       | 40        |
| 7.5 Standard Extended / Specialized / Private SOP Classes . . . . .             | 40        |
| 7.6 Private Transfer Syntaxes . . . . .                                         | 40        |

## 2 Introduction

This DICOM Conformance Statement specifies the behavior and functionality of the OCULUS Smartfield system, with regard to supported DICOM networking SOP Classes. OCULUS Smartfield is intended for testing the visual field of the human eye.

### 2.1 Revision History

| Document Version | Date of Issue | Author      | Description                                          |
|------------------|---------------|-------------|------------------------------------------------------|
| Version 1.76     | 2021-10-04    | Daniel Ache | Initial release of this document                     |
| Version 2.0      | 2019-03-25    | Daniel Ache | Added "Ophthalmic Visual Field Static Perimetry IOD" |
| Version 2.1      | 2023-06-19    | Daniel Ache | rebuild document in L <sup>A</sup> T <sub>E</sub> X  |

### 2.2 Audience

This document is written for the people that need to understand how OCULUS Smartfield will integrate into their healthcare facility. This includes both those responsible for overall imaging network policy and architecture, as well as integrators who need to have a detailed understanding of the DICOM features of the product. This document contains some basic DICOM definitions so that any reader may understand how this product implements DICOM features. However, integrators are expected to fully understand all the DICOM terminology, how the tables in this document relate to the product's functionality, and how that functionality integrates with other devices that support compatible DICOM features.

### 2.3 Remarks

The scope of this DICOM Conformance Statement is to facilitate integration between OCULUS Smartfield and other DICOM products. The Conformance Statement should be read and understood in conjunction with the DICOM Standard. DICOM by itself does not guarantee interoperability. The Conformance Statement does, however, facilitate a first-level comparison for interoperability between different applications supporting compatible DICOM functionality. This Conformance Statement is not supposed to replace validation with other DICOM equipment to ensure proper exchange of intended information. In fact, the user should be aware of the following important issues:

- The comparison of different Conformance Statements is just the first step towards assessing interconnectivity and interoperability between the product and other DICOM conformant equipment.
- Test procedures should be defined and executed to validate the required level of interoperability with specific compatible DICOM equipment, as established by the healthcare facility.

## 2.4 Terms and Definitions

Informal definitions are provided for the following terms used in this Conformance Statement. The DICOM Standard [DICOM] is the authoritative source for formal definitions of these terms.

- **Abstract Syntax:** the information agreed to be exchanged between applications, generally equivalent to a Service/Object Pair (SOP) Class. Examples: Verification SOP Class, Modality Worklist Information Model Find SOP Class, Computed Radiography Image Storage SOP Class.
- **Application Entity (AE):** an end point of a DICOM information exchange, including the DICOM network or media interface software; i. e., the software that sends or receives DICOM information objects or messages. A single device may have multiple Application Entities.
- **Application Entity Title:** the externally known name of an Application Entity, used to identify a DICOM application to other DICOM applications on the network.
- **Application Context:** the specification of the type of communication used between Application Entities. Example: DICOM network protocol.
- **Association:** a network communication channel set up between Application Entities.
- **Attribute:** a unit of information in an object definition; a data element identified by a tag. The information may be a complex data structure (Sequence), itself composed of lower level data elements. Examples: Patient ID (0010,0020), Accession Number (0008,0050), Photometric Interpretation (0028,0004), Procedure Code Sequence (0008,1032).
- **Information Object Definition (IOD):** the specified set of Attributes that comprise a type of data object; does not represent a specific instance of the data object, but rather a class of similar data objects that have the same properties. The Attributes may be specified as Mandatory (Type 1), Required but possibly unknown (Type 2), or Optional (Type 3), and there may be conditions associated with the use of an Attribute (Types 1C and 2C). Examples: MR Image IOD, CT Image IOD, Print Job IOD.
- **Joint Photographic Experts Group (JPEG):** a set of standardized image compression techniques, available for use by DICOM applications.
- **Media Application Profile:** the specification of DICOM information objects and encoding exchanged on removable media (e. g., CDs).
- **Module:** a set of Attributes within an Information Object Definition that are logically related to each other. Example: Patient Module includes Patient Name, Patient ID, Patient Birth Date, and Patient Sex.
- **Negotiation:** first phase of Association establishment that allows Application Entities to agree on the types of data to be exchanged and how that data will be encoded. Presentation Context: the set of DICOM network services used over an Association, as negotiated between Application Entities; includes Abstract Syntaxes and Transfer Syntaxes.
- **Protocol Data Unit (PDU):** a packet (piece) of a DICOM message sent across the network. Devices must specify the maximum size packet they can receive for DICOM messages.
- **Service Class Provider (SCP):** role of an Application Entity that provides a DICOM network service; typically, a server that performs operations requested by another Application Entity (Service Class User). Examples: Picture Archiving and Communication System (image storage SCP, and image query/retrieve SCP), Radiology Information System (modality worklist SCP).
- **Service Class User (SCU):** role of an Application Entity that uses a DICOM network service; typically, a client. Examples: imaging modality (image storage SCU, and modality worklist SCU), imaging workstation (image query/retrieve SCU)

- **Service/Object Pair (SOP) Class:** the specification of the network or media transfer (service) of a particular type of data (object); the fundamental unit of DICOM interoperability specification. Examples: Ultrasound Image Storage Service, Basic Grayscale Print Management. Service/Object Pair (SOP) Instance: an information object; a specific occurrence of information exchanged in a SOP Class. Examples: a specific X-Ray image.
- **Tag:** a 32-bit identifier for a data element, represented as a pair of four digit hexadecimal numbers, the “group” and the “element”. If the “group” number is odd, the tag is for a private (manufacturer-specific) data element. Examples: (0010,0020) [Patient ID], (07FE,0010) [Pixel Data], (0019,0210) [private data element]
- **Transfer Syntax:** the encoding used for exchange of DICOM information objects and messages. Examples: JPEG compressed (images), little endian explicit value representation.
- **Unique Identifier (UID):** a globally unique “dotted decimal” string that identifies a specific object or a class of objects; an ISO-8824 Object Identifier. Examples: Study Instance UID, SOP Class UID, SOP Instance UID.
- **Value Representation (VR):** the format type of an individual DICOM data element, such as text, an integer, a person’s name, or a code. DICOM information objects can be transmitted with either explicit identification of the type of each data element (Explicit VR), or without explicit identification (Implicit VR); with Implicit VR, the receiving application must use a DICOM data dictionary to look up the format of each data element.

## 2.5 Basics of DICOM Communication

This section describes terminology used in this Conformance Statement for the non-specialist. The key terms used in the Conformance Statement are highlighted in italics below. This section is not a substitute for training about DICOM, and it makes many simplifications about the meanings of DICOM terms.

Two Application Entities (devices) that want to communicate with each other over a network using DICOM protocol must first agree on several things during an initial network “handshake”. One of the two devices must initiate an Association (a connection to the other device), and ask if specific services, information, and encoding can be supported by the other device (Negotiation).

DICOM specifies a number of network services and types of information objects, each of which is called an Abstract Syntax for the Negotiation. DICOM also specifies a variety of methods for encoding data, denoted Transfer Syntaxes. The Negotiation allows the initiating Application Entity to propose combinations of Abstract Syntax and Transfer Syntax to be used on the Association; these combinations are called Presentation Contexts. The receiving Application Entity accepts the Presentation Contexts it supports.

For each Presentation Context, the Association Negotiation also allows the devices to agree on Roles - which one is the Service Class User (SCU - client) and which is the Service Class Provider (SCP - server). Normally the device initiating the connection is the SCU, i. e., the client system calls the server, but not always.

The Association Negotiation finally enables exchange of maximum network packet (PDU) size, security information, and network service options (called Extended Negotiation information).

The Application Entities, having negotiated the Association parameters, may now commence exchanging data. Common data exchanges include queries for worklists and lists of stored images, transfer of image objects and analyses (structured reports), and sending images to film printers. Each exchangeable unit of data is formatted by the sender in accordance with the appropriate Information Object Definition, and sent using the negotiated Transfer Syntax. There is a Default Transfer Syntax that all systems must accept, but it may not be the most efficient for some use cases. Each transfer is explicitly acknowledged by the receiver with a Response Status indicating success, failure, or that query or retrieve operations are still in process.

Two Application Entities may also communicate with each other by exchanging media (such as a CD-R). Since there is no Association Negotiation possible, they both use a Media Application Profile that specifies “pre-negotiated” exchange media format, Abstract Syntax, and Transfer Syntax.

## 2.6 Abbreviations

The following list illustrates all abbreviations that are used in this document.

- **ACSE:** Association Control Service Element
- **AE:** Application Entity
- **CD:** Compact Disc
- **CD-R:** Compact Disc – Recordable
- **CT:** Computer Tomography
- **DICOM:** Digital Imaging and Communications in Medicine
- **DIMSE:** DICOM Message Service Element
- **ID:** Identification / Identifier
- **IOD:** Information Object Definition
- **IPv4:** Internet Protocol version 4
- **IPv6:** Internet Protocol version 6
- **ISO:** International Standards Organization
- **JPEG:** Joint Photographic Experts Group
- **MR:** Magnetic Resonance
- **MWL:** Modality Worklist Management
- **PDF:** Portable Document Format
- **PDU:** Protocol Data Unit
- **SCP:** Service Class Provider
- **SCU:** Service Class User
- **SOP:** Service Object Pair
- **TCP/IP:** Transmission Control Protocol / Internet Protocol
- **UID:** Unique Identifier
- **VM:** Value Multiplicity
- **VR:** Value Representation

### 2.6.1 References

[DICOM] NEMA Standards Publication PS 3.1-18, Digital Imaging and Communications in Medicine  
[DICOM], 2022, available at  
<https://dicomstandard.org/>





### 3.1.3 Sequencing of Real-World Activities

All SCU activities are initiated in the user interface. For each remote AE, a new background is started which allows for multiple associations at the same time.

## 3.2 AE Specifications

### 3.2.1 WORKLIST-SCU

#### 3.2.1.1 SOP Classes

WORKLIST-SCU provides standard conformance to the following DICOM SOP classes.

Table 3.1: SOP Classes for AE WORKLIST-SCU

| SOP Class Name                             | SOP Class UID          | SCU | SCP |
|--------------------------------------------|------------------------|-----|-----|
| Modality Worklist Information Model - FIND | 1.2.840.10008.5.1.4.31 | Yes | No  |

#### 3.2.1.2 Association Policies

**3.2.1.2.1 General** WORKLIST-SCU initiates but never accepts associations. The DICOM standard application context name, which is always proposed, is:

Table 3.2: DICOM Application Context for AE WORKLIST-SCU

|                          |                       |
|--------------------------|-----------------------|
| Application context name | 1.2.840.10008.3.1.1.1 |
|--------------------------|-----------------------|

The maximum PDU size can be configured at installation time in the range 4096 to 131072 bytes.

The default is 16384 bytes.

SOP Class extended negotiation is not supported.

#### 3.2.1.2.2 Number of Associations

Table 3.3: Number of Associations as an Association Initiator for AE WORKLIST-SCU

|                                             |   |
|---------------------------------------------|---|
| Maximum number of simultaneous associations | 1 |
|---------------------------------------------|---|

**3.2.1.2.3 Asynchronous Nature** Asynchronous mode of operation is not supported.

#### 3.2.1.2.4 Implementation Identifying Information

Table 3.4: DICOM Implementation Class and Version for AE WORKLIST-SCU

|                             |                             |
|-----------------------------|-----------------------------|
| Implementation Class UID    | 1.2.276.0.7230010.3.0.3.5.5 |
| Implementation Version Name | OFFIS_DCMTK_366             |

#### 3.2.1.3 Association Initiation Policy

WORKLIST-SCU attempts to initiate a new association for the Modality Worklist Information Model - FIND SOP Class each time the user opens the patient list and the system wants to send a query.

### 3.2.1.3.1 Activity - System sends a query and shows matches

#### Description and Sequencing of Activities

WORKLIST-SCU sends a query to a selected remote AE and shows the matching worklist entries (if any) in the patient list. For each query to be performed, a single attempt will be made. If the send fails, no retry will be performed but an error message will be reported to the user.

#### Proposed Presentation Contexts

Table 3.5: Proposed Presentation Contexts for AE WORKLIST-SCU and Real-World Activity 'System Sends a query and Shows Matches'

| Presentation Context Table                 |                        |                                                                              |                                                             |      |           |
|--------------------------------------------|------------------------|------------------------------------------------------------------------------|-------------------------------------------------------------|------|-----------|
| Abstract Syntax                            |                        | Transfer Syntax                                                              |                                                             | Role | Ext. Neg. |
| Name                                       | UID                    | Name List                                                                    | UID List                                                    |      |           |
| Modality Worklist Information Model - FIND | 1.2.840.10008.5.1.4.31 | Implicit VR Little Endian, Explicit VR Little Endian, Explicit VR Big Endian | 1.2.840.10008.1.2, 1.2.840.10008.1.2.1, 1.2.840.10008.1.2.2 | SCU  | None      |

#### Extended Negotiation

No extended negotiation is performed.

#### SOP Specific Conformance to Modality Worklist SOP Class

SOP Specific Conformance to Modality Worklist SOP Class WORKLIST-SCU provides standard conformance to the Modality Worklist Information Model - FIND SOP Class as an SCU.

Unexpected attributes returned in a C-FIND response are ignored in the same manner as requested (optional) return attributes not returned by the SCP. Non-matching responses returned by the SCP due to unsupported matching keys are not filtered locally by the WORKLIST-SCU and thus will still be shown in the patient list. No attempt is made to filter out duplicate responses.

No cancel requests are ever issued.

Specific Character Set (0008,0005) is included in the query. If present in the response, its element value will be used to identify character sets other than the default character set for display. Please note, however, that only ISO\_IR 100 (ISO 8859-1, Latin 1) or ISO\_IR 192 (Utf-8) are supported (see sect. 4, page 17). If another specific character set is specified, the worklist entry will not be shown to the user but a message will be reported to the service logs.

The WORKLIST-SCU application entity uses the following attributes from the Modality Worklist Information Model.

Table 3.6: Request Identifiers for AE WORKLIST-SCU

| Attribute                                 | Tag         | Types of Matching     |
|-------------------------------------------|-------------|-----------------------|
| Accession Number                          | {0008,0050} | Universal Matching    |
| Referring Physician's Name                | {0008,0090} | Universal Matching    |
| Patient's Name                            | {0010,0010} | Universal Matching    |
| Patient ID                                | {0010,0020} | Universal Matching    |
| Patient's Birth Date                      | {0010,0030} | Universal Matching    |
| Patient's Sex                             | {0010,0040} | Universal Matching    |
| Study Instance UID                        | {0020,000D} | Universal Matching    |
| Requesting Physician                      | {0032,1032} | Universal Matching    |
| Requested Procedure Description           | {0032,1060} | Universal Matching    |
| Scheduled Procedure Step Sequence         | {0040,0100} | Universal Matching    |
| >Modality                                 | {0008,0060} | Universal Matching    |
| >Requested Contrast Agent                 | {0032,1070} | Universal Matching    |
| >Scheduled Station AE Title               | {0040,0001} | Single Value Matching |
| >Scheduled Procedure Step Start Date      | {0040,0002} | Single Value Matching |
| >Scheduled Procedure Step Start Time      | {0040,0003} | Universal Matching    |
| >Scheduled Performing Physician's Name    | {0040,0006} | Universal Matching    |
| >Scheduled Procedure Step Description     | {0040,0007} | Universal Matching    |
| >Scheduled Procedure Step ID              | {0040,0009} | Universal Matching    |
| >Scheduled Station Name                   | {0040,0010} | Universal Matching    |
| >Scheduled Procedure Step Location        | {0040,0011} | Universal Matching    |
| >Pre-Medication                           | {0040,0012} | Universal Matching    |
| >Comments On The Scheduled Procedure Step | {0040,0400} | Universal Matching    |
| Performed Procedure Step ID               | {0040,0253} | Universal Matching    |

Further attributes can be added by changing the configuration of WORKLIST- SCU.

#### Presentation Context Acceptance Criterion

WORKLIST-SCU does not accept associations.

#### Transfer Syntax Selection Policies

If offered a choice of transfer syntaxes in the accepted presentation contexts, the default behavior of WORKLIST-SCU is to prefer the explicit VR transfer syntax with the byte order matching the local machine byte order, i.e. little endian on a Windows PC. If this transfer syntax is not available, the explicit VR transfer syntax with opposite byte order is selected. If this is also unavailable, implicit VR little endian is selected. Finally, any matching presentation context is accepted independent of the transfer syntax.

The default behavior may be modified by changing the configuration of WORKLIST-SCU.

#### Response Status

WORKLIST-SCU will behave as described in the table below when receiving the C-FIND response command message.

Table 3.7: Response Status Handling Behavior for AE WORKLIST-SCU and Real-World Activity 'System Sends a Query and Shows Matches'

| Service Status | Further Meaning                                                                                                                      | Error Code | Behavior                                                                                              |
|----------------|--------------------------------------------------------------------------------------------------------------------------------------|------------|-------------------------------------------------------------------------------------------------------|
| Refused        | Out of resources                                                                                                                     | A7xx       | This is treated as a permanent failure.                                                               |
| Error          | Identifier does not match SOP class                                                                                                  | A9xx       | This is treated as a permanent failure.                                                               |
|                | Unable to process                                                                                                                    | Cxxx       | This is treated as a permanent failure.                                                               |
| Cancel         | Matching terminated due to cancel request                                                                                            | FE00       | Should never occur since cancel requests are never issued.                                            |
| Success        | Matching is complete - No final identifier is supplied                                                                               | 0000       | Current query is completed successfully; resulting list of worklist entries is presented to the user. |
| Pending        | Matches are continuing - Current match is supplied and any optional keys were supported in the same manner as required keys          | FF00       | Current match is added to the internal list of worklist entries but not yet presented to the user.    |
|                | Matches are continuing - Warning that one or more optional keys were not supported for existence and/or matching for this identifier | FF01       | Current match is added to the internal list of worklist entries but not yet presented to the user.    |

Table 3.8: Response Status Handling Behavior for AE WORKLIST-SCU and Real-World Activity 'System Sends a Query and Shows Matches'

| Exception                                                                                                                              | Behavior                                                                         |
|----------------------------------------------------------------------------------------------------------------------------------------|----------------------------------------------------------------------------------|
| TCP/IP connection could not be established within the specified time range (configurable parameter).                                   | An error message is reported to both the service logs and to the user interface. |
| ASCE response message could not be received within the specified time range (configurable parameter).                                  | An error message is reported to both the service logs and to the user interface. |
| DIMSE response message could not be received within the specified time range (configurable parameter).                                 | An error message is reported to both the service logs and to the user interface. |
| Association aborted by the SCP using A-ABORT or the network layers indicate communication loss (i. e. low-level TCP/IP socket closure) | An error message is reported to both the service logs and to the user interface. |

#### 3.2.1.4 Association Acceptance Policy

WORKLIST-SCU does not accept associations.

## 3.2.2 STORAGE-SCU

### 3.2.2.1 SOP Classes

STORAGE-SCU provides standard conformance to the following DICOM SOP classes.

Table 3.9: SOP Classes for AE STORAGE-SCU

| SOP Class Name                                                | SOP Class UID                 | SCU | SCP |
|---------------------------------------------------------------|-------------------------------|-----|-----|
| Encapsulated PDF Storage                                      | 1.2.840.10008.5.1.4.1.1.104.1 | Yes | No  |
| Secondary Capture Image Storage                               | 1.2.840.10008.5.1.4.1.1.7     | Yes | No  |
| Ophthalmic Visual Field Static Perimetry Measurements Storage | 1.2.840.10008.5.1.4.1.1.80.1  | Yes | No  |

### 3.2.2.2 Association Policies

#### 3.2.2.2.1 General

STORAGE-SCU initiates but never accepts associations. The DICOM standard application context name, which is always proposed, is:

Table 3.10: DICOM Application Context for AE STORAGE-SCU

|                          |                       |
|--------------------------|-----------------------|
| Application context name | 1.2.840.10008.3.1.1.1 |
|--------------------------|-----------------------|

The maximum PDU size can be configured at installation time in the range 4096 to 131072 bytes. The default is 16384 bytes. SOP Class extended negotiation is not supported.

#### 3.2.2.2.2 Number of Associations

Table 3.11: Number of Associations as an Associations Initiator for AE STORAGE-SCU

|                                             |   |
|---------------------------------------------|---|
| Maximum number of simultaneous associations | 1 |
|---------------------------------------------|---|

#### 3.2.2.2.3 Asynchronous Nature

Asynchronous mode of operation is not supported.

#### 3.2.2.2.4 Implementation Identifying Information

Table 3.12: DICOM Implementation Class and Version for AE STORAGE-SCU

|                             |                             |
|-----------------------------|-----------------------------|
| Implementation Class UID    | 1.2.276.0.7230010.3.0.3.5.5 |
| Implementation Version Name | OFFIS_DCMTK_366             |

### 3.2.2.3 Association Initiation Policy

STORAGE-SCU attempts to initiate a new association each time the system creates an instance and wants to send it.

#### 3.2.2.3.1 Activity - System creates new instance and sends it

### Description and Sequencing of Activities

STORAGE-SCU sends a newly created instance to a selected remote AE. For each instance to be transferred, a single attempt will be made. If the send fails, no retry will be performed but an error message will be reported to the service logs.

### Proposed Presentation Contexts

Table 3.13: Proposed Presentation Contexts for AE STORAGE-SCU and Real-World Activity 'System Creates New Instance and Sends It'

| Presentation Context Table                                    |                               |                                                                                                           |                                                                                              |      |      |
|---------------------------------------------------------------|-------------------------------|-----------------------------------------------------------------------------------------------------------|----------------------------------------------------------------------------------------------|------|------|
| Abstract Syntax                                               |                               | Transfer Syntax                                                                                           |                                                                                              | Role | Ext. |
| Name                                                          | UID                           | Name List                                                                                                 | UID List                                                                                     |      | Neg. |
| Encapsulated PDF Storage                                      | 1.2.840.10008.5.1.4.1.1.104.1 | Implicit VR Little Endian,<br>Explicit VR Little Endian,<br>Explicit VR Big Endian                        | 1.2.840.10008.1.2,<br>1.2.840.10008.1.2.1,<br>1.2.840.10008.1.2.2                            | SCU  | None |
| Secondary Capture Image Storage                               | 1.2.840.10008.5.1.4.1.1.7     | Implicit VR Little Endian,<br>Explicit VR Little Endian,<br>Explicit VR Big Endian,<br>JPEG Baseline (P1) | 1.2.840.10008.1.2,<br>1.2.840.10008.1.2.1,<br>1.2.840.10008.1.2.2,<br>1.2.840.10008.1.2.4.50 | SCU  | None |
| Ophthalmic Visual Field Static Perimetry Measurements Storage | 1.2.840.10008.5.1.4.1.1.80.1  | Implicit VR Little Endian,<br>Explicit VR Little Endian,<br>Explicit VR Big Endian                        | 1.2.840.10008.1.2,<br>1.2.840.10008.1.2,<br>1.2.840.10008.1.2.1,<br>1.2.840.10008.1.2.2      | SCU  | None |

### Extended Negotiation

No extended negotiation is performed.

### SOP Specific Conformance to Storage SOP Classes

STORAGE-SCU provides standard conformance to the Storage Service Class as an SCU.

### Presentation Context Acceptance Criterion

STORAGE-SCU does not accept associations.

### Transfer Syntax Selection Policies

If offered a choice of transfer syntaxes in the accepted presentation contexts, the default behavior of STORAGE-SCU is to prefer the transfer syntax that is used for the DICOM file stored on the local system. If this transfer syntax is not available, an explicit VR uncompressed transfer syntax is selected. If this is also unavailable, implicit VR little endian is selected. Finally, any matching presentation context is accepted independent of the transfer syntax.

The default behavior may be modified by changing the configuration of STORAGE-SCU.

### Response Status

STORAGE-SCU will behave as described in the table below when receiving the C-STORE response command message.

Table 3.14: Response Status Handling Behavior for AE STORAGE-SCU and Real-World Activity 'System Creates New Instance and Sends It'

| Service Status | Further Meaning                   | Error Code | Behavior                                                                                                                                                  |
|----------------|-----------------------------------|------------|-----------------------------------------------------------------------------------------------------------------------------------------------------------|
| Refused        | Out of resources                  | A7xx       | This is treated as a permanent failure. An error message is reported to the service logs. No message is posted to the user interface.                     |
| Error          | Data set does not match SOP class | A9xx       | This is treated as a permanent failure. An error message is reported to the service logs. No message is posted to the user interface.                     |
|                | Cannot understand                 | Cxxx       | This is treated as a permanent failure. An error message is reported to the service logs. No message is posted to the user interface.                     |
| Warning        |                                   | Bxxx       | Transmission of the DICOM instance is considered successful. A warning message is output to the service logs. No message is posted to the user interface. |
| Success        |                                   | 0000       | The DICOM instance was successfully received by the remote AE. No message is reported to the service logs or posted to the user interface.                |

Draft



Table 3.15: Communication Failure Behavior for AE STORAGE-SCU and Real-World Activity 'System Creates New Instance and Sends It'

| Exception                                                                                                                              | Behavior                                                                                      |
|----------------------------------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------------------|
| TCP/IP connection could not be established within the specified time range (configurable parameter).                                   | An error message is reported to the service logs. No message is posted to the user interface. |
| ASCE response message could not be received within the specified time range (configurable parameter).                                  | An error message is reported to the service logs. No message is posted to the user interface. |
| DIMSE response message could not be received within the specified time range (configurable parameter).                                 | An error message is reported to the service logs. No message is posted to the user interface. |
| Association aborted by the SCP using A-ABORT or the network layers indicate communication loss (i. e. low-level TCP/IP socket closure) | An error message is reported to the service logs. No message is posted to the user interface. |

#### 3.2.2.4 Association Acceptance Policy

STORAGE-SCU does not accept associations.

### 3.3 Physical Network Interfaces

#### 3.3.1 Physical Network Interface

The DICOM applications of OCULUS Smartfield are indifferent to the physical medium over which TCP/IP is used.

#### 3.3.2 Additional Protocols

When host names rather than IP addresses are used in the configuration to specify presentation addresses for remote AEs, the application is dependent on the name resolution mechanism of the underlying operating system.

#### 3.3.3 IPv4 and IPv6 Support

OCULUS Smartfield only supports IPv4 connections.

### 3.4 Configuration

The configuration can be changed by the user in the device DICOM settings (DICOM.ini). Details are described in the document "Configuration of the OCULUS DICOM-interface".

#### 3.4.1 AE Title / Presentation Address Mapping

The Calling AE Title of the local AEs is configurable in the preferences dialog and the device's configuration file ("dicom.ini"). The mapping of the logical name by which remote AEs are described in the user interface to Called AE Titles as well as presentation address (hostname or IP address and port number) is configurable in the preferences dialog and the configuration file.

### 3.4.1.1 Local AE Titles

Table 3.16: AE Title Configuration Table

| Application Entity | Default AE Title | Default TCP/IP Port |
|--------------------|------------------|---------------------|
| WORKLIST-SCU       | Smartfield       | -                   |
| STORAGE-SCU        | Smartfield       | -                   |

### 3.4.1.2 Remote AE Titles

The AE Title, host name or IP address, and port number of remote AEs are configured in the preferences dialog.

## 3.4.2 Configurable Parameters

Concerning the WORKLIST-SCU application entity, the following parameters are configurable.

Table 3.17: Configurable Parameters for AE WORKLIST-SCU

| Parameter                                               | Configurable | Default Value |
|---------------------------------------------------------|--------------|---------------|
| Time-out waiting for response to TCP/IP connect request | Yes          | unlimited     |
| Time-out for ACSE messages                              | Yes          | 30 seconds    |
| Time-out for DIMSE messages                             | Yes          | unlimited     |
| Maximum PDU size the AE can receive                     | Yes          | 16384 bytes   |

Concerning the STORAGE-SCU application entity, the following parameters are configurable.

Table 3.18: Configurable Parameters for AE STORAGE-SCU

| Parameter                                               | Configurable | Default Value |
|---------------------------------------------------------|--------------|---------------|
| Time-out waiting for response to TCP/IP connect request | Yes          | unlimited     |
| Time-out for ACSE messages                              | Yes          | 30 seconds    |
| Time-out for DIMSE messages                             | Yes          | unlimited     |
| Maximum PDU size the AE can receive                     | Yes          | 16384 bytes   |

## 4 Media Storage

OCULUS Smartfield does not support any media services.

Draft

## 5 Support of extended character sets

OCULUS Smartfield supports the ISO\_IR 100 (ISO 8859-1, Latin 1) or ISO\_IR 192 (Utf-8) extended character sets for the WORKLIST-SCU and for the created SOP instances.

Draft

## 6 Security

### 6.1 Security Profiles

OCULUS Smartfield does not support any security profiles.

### 6.2 Association Level Security

OCULUS Smartfield does not support any association level security.

### 6.3 Application Level Security

OCULUS Smartfield does not support any application level security.

Draft

## 7 Annexes

### 7.1 IOD Contents

#### 7.1.1 Created SOP Instances

##### 7.1.1.1 Encapsulated PDF IOD

OCULUS Smartfield creates Encapsulated PDF IOD objects containing a printout of the current examination. The following tables describe the modules and attributes of the underlying IOD containing a printout image of the examination. Most attributes that are never present in a created SOP instance are omitted from the tables in order to increase the readability.

Table 7.1: Encapsulated PDF IOD Modules

| IE                    | Module                       | Reference  | Presence of Module |
|-----------------------|------------------------------|------------|--------------------|
| Patient               | Patient                      | Table 7.9  | Always             |
|                       | Clinical Trial Subject       | -          | Never              |
| Study                 | General Study                | Table 7.10 | Always             |
|                       | Patient Study                | -          | Never              |
|                       | Clinical Trial Subject       | -          | Never              |
| Series                | Encapsulated Document Series | Table 7.14 | Always             |
|                       | Clinical Trial Subject       | -          | Never              |
| Equipment             | General Equipment            | Table 7.11 | Always             |
|                       | SC Equipment                 | Table 7.12 | Always             |
| Encapsulated Document | Encapsulated Document        | Table 7.15 | Always             |
|                       | SOP Common                   | Table 7.16 | Always             |
|                       | Private Data                 | Table 7.13 | Not always         |

### 7.1.1.2 Secondary Capture Image IOD

OCULUS Smartfield creates Secondary Capture Image IOD objects containing a printout of the current examination. The following tables describe the modules and attributes of the underlying IOD. Most attributes that are never present in a created SOP instance are omitted from the tables in order to increase the readability.

Table 7.2: Secondary Capture Image IOD Modules

| IE        | Module                 | Reference  | Presence of Module |
|-----------|------------------------|------------|--------------------|
| Patient   | Patient                | Table 7.9  | Always             |
|           | Clinical Trial Subject | -          | Never              |
| Study     | General Study          | Table 7.10 | Always             |
|           | Patient Study          | -          | Never              |
|           | Clinical Trial Subject | -          | Never              |
| Series    | General Series         | Table 7.21 | Always             |
|           | Clinical Trial Subject | -          | Never              |
| Equipment | General Equipment      | Table 7.11 | Always             |
|           | SC Equipment           | Table 7.12 | Always             |
| Image     | General Image          | Table 7.18 | Always             |
|           | Image Pixel            | Table 7.19 | Always             |
|           | Device                 | -          | Never              |
|           | Specimen               | -          | Never              |
|           | SC Image               | -          | Never              |
|           | Overlay Plane          | -          | Never              |
|           | Modality LUT           | -          | Never              |
|           | VOI LUT                | -          | Never              |
|           | SOP Common             | Table 7.20 | Always             |
|           | Private Data           | Table 7.13 | Not always         |

### 7.1.1.3 Ophthalmic Visual Field Static Perimetry Measurements IOD

OCULUS Smartfield creates Ophthalmic Visual Field Static Perimetry Measurements IOD objects. The following tables describe the modules and attributes of the underlying IOD. Most attributes that are never present in a created SOP instance are omitted from the tables in order to increase the readability.

Table 7.3: Ophthalmic Visual Field Static Perimetry IOD Modules

| IE           | Module                                                           | Reference  | Presence of Module |
|--------------|------------------------------------------------------------------|------------|--------------------|
| Patient      | Patient                                                          | Table 7.9  | Always             |
|              | Clinical Trial Subject                                           | -          | Never              |
| Study        | General Study                                                    | Table 7.10 | Always             |
|              | Patient Study                                                    | -          | Never              |
|              | Clinical Trial Subject                                           | -          | Never              |
| Series       | General Series                                                   | Table 7.21 | Always             |
|              | Clinical Trial Series                                            | -          | Never              |
|              | Visual Field Static Perimetry Measurements Series                | Table 7.21 | Always             |
| Equipment    | General Equipment                                                | Table 7.11 | Always             |
|              | Enhanced General Equipment                                       | Table 7.11 | Always             |
| Measurements | Visual Field Static Perimetry Test Parameters                    | Table 7.4  | Always             |
|              | Visual Field Static Perimetry Test Reliability                   | Table 7.5  | Always             |
|              | Visual Field Static Perimetry Test Measurements                  | Table 7.6  | Always             |
|              | Visual Field Static Perimetry Test Results                       | Table 7.7  | Always             |
|              | Ophthalmic Patient Clinical Information and Test Lens Parameters | Table 7.8  | Always             |
|              | SOP Common                                                       | Table 7.20 | Always             |
|              | Private Data                                                     | Table 7.13 | Not always         |



Table 7.4: Visual Field Static Perimetry Test Parameters

| Attribute Name                 | Tag         | VR | Value      | Presence of Value |
|--------------------------------|-------------|----|------------|-------------------|
| Visual Field Horizontal Extent | [0024,0010] | FL |            | Always            |
| Visual Field Vertical Extent   | [0024,0011] | FL |            | Always            |
| Visual Field Shape             | [0024,0012] | CS | CIRCLE     | Always            |
| Maximum Stimulus Luminance     | [0024,0018] | FL |            | Always            |
| Background Luminance           | [0024,0020] | FL |            | Always            |
| Stimulus Color Code Sequence   | [0024,0021] | SQ | Table 7.26 | Always            |
| Background Illumination Color  | [0024,0024] | SQ | Table 7.26 | Always            |
| Code Sequence Stimulus Area    | [0024,0025] | FL |            | Always            |
| Stimulus Presentation Time     | [0024,0028] | FL |            | Always            |

Table 7.5: Visual Field Static Perimetry Test Reliability Module

| Attribute Name                          | Tag         | VR | Value      | Presence of Value    |
|-----------------------------------------|-------------|----|------------|----------------------|
| Fixation Sequence                       | [0024,0032] | SQ |            | Always               |
| > Fixation Monitoring Code Sequence     | [0024,0033] | SQ | Table 7.25 | Always               |
| > Fixation Checked Quantity             | [0024,0035] | US |            | Always               |
| > Patient Not Properly Fixated Quantity | [0024,0036] | US |            | Always               |
| > Excessive Fixation Losses Data Flag   | [0024,0039] | CS |            | Always               |
| > Excessive Fixation Losses             | [0024,0040] | CS |            | Not always present   |
| Visual Field Catch Trial Sequence       | [0024,0034] | SQ |            | Always               |
| > Catch Trials Data Flag                | [0024,0055] | CS | "YES"/"NO" | Always               |
| > Negative Catch Trials Quantity        | [0024,0048] | US |            | Not always present   |
| > False Negatives Quantity              | [0024,0050] | US |            | Not always present   |
| > False Negatives Estimate Flag         | [0024,0045] | CS | "YES"/"NO" | Always               |
| > False Negatives Estimate              | [0024,0046] | FL |            | Always               |
| > Excessive False Negatives Data Flag   | [0024,0051] | CS | "YES"/"NO" | Always               |
| > Excessive False Negatives             | [0024,0052] | CS | "YES"      | "Not always present" |
| > Positive Catch Trials Quantity        | [0024,0056] | US |            | Not always present   |
| > False Positives Quantity              | [0024,0060] | US |            | Not always present   |
| > False Positives Estimate Flag         | [0024,0053] | CS | "YES"/"NO" | Always               |
| > False Positives Estimate              | [0024,0054] | FL |            | Not always present   |
| > Excessive False Positives Data Flag   | [0024,0061] | CS | "YES"/"NO" | Always               |
| > Excessive False Positives             | [0024,0062] | CS | "YES"      | Not always present   |

Table 7.6: Visual Field Static Perimetry Test Measurement

| Attribute Name                                              | Tag         | VR | Value                         | Presence of Value  |
|-------------------------------------------------------------|-------------|----|-------------------------------|--------------------|
| Measurement Laterality                                      | (0024,0113) | CS | "L", "R", "B"                 | Always             |
| Presented Visual Stimuli Data Flag                          | (0024,0037) | CS | "YES"                         | Always             |
| Number Of Visual Stimuli                                    | (0024,0038) | US |                               | Always             |
| Visual Field Test Duration                                  | (0024,0088) | FL |                               | Always             |
| Foveal Sensitivity Measured                                 | (0024,0086) | CS | "YES"/"NO"                    | Always             |
| Foveal Sensitivity                                          | (0024,0087) | FL |                               | Not always present |
| Foveal Point Normative Data Flag                            | (0024,0117) | CS | "YES"/"NO"                    | Always             |
| Foveal Point Probability Value                              | (0024,0118) | FL |                               | Not always present |
| Screening Baseline Measured                                 | (0024,0120) | CS | "NO"                          | Always             |
| Blind Spot Localized                                        | (0024,0106) | CS | "YES"/"NO"                    | Always             |
| Blind Spot XCoordinate                                      | (0024,0107) | FL |                               | Not always present |
| Blind Spot YCoordinate                                      | (0024,0108) | FL |                               | Not always present |
| Minimum Sensitivity Value                                   | (0024,0105) | FL |                               | Always             |
| Test Point Normals Data Flag                                | (0024,0057) | CS | "YES"/"NO"                    | Always             |
| Test Point Normals Sequence                                 | (0024,0058) | SQ |                               | Not always present |
| > Data Set Name                                             | (0024,0306) | LO | "OCULUS normative data"       | Not always present |
| > Data Set Version                                          | (0024,0307) | LO |                               | Not always present |
| > Data Set Source                                           | (0024,0308) | LO | "OCU"                         | Not always present |
| Age Corrected Sensitivity Deviation Algorithm Sequence      | (0024,0065) | SQ |                               | Not always present |
| > Algorithm Family Code Sequence                            | (0066,002F) | SQ |                               | Not always present |
| > > Code Value                                              | (0008,0100) | SH | "PERIMETRY"                   | Always             |
| > > Code Scheme Designator                                  | (0008,0102) | SH | "99OCU_PERIMETRY"             | Always             |
| > > Code Meaning                                            | (0008,0104) | LO | "OCULUS Perimetry Algorithms" | Always             |
| > Algorithm Version                                         | (0066,0031) | LO |                               | Always             |
| > Algorithm Name                                            | (0066,0036) | LO |                               | Always             |
| Generalized Defect Sensitivity Deviation Algorithm Sequence | (0024,0067) | SQ |                               | Not always present |
| > Algorithm Family Code Sequence                            | (0066,002F) | SQ |                               | Not always present |
| > > Code Value                                              | (0008,0100) | SH | "PERIMETRY"                   | Always             |
| > > Code Scheme Designator                                  | (0008,0102) | SH | "99OCU_PERIMETRY"             | Always             |
| > > Code Meaning                                            | (0008,0104) | LO | "OCULUS Perimetry Algorithms" | Always             |
| > Algorithm Version                                         | (0066,0031) | LO |                               | Always             |
| > Algorithm Name                                            | (0066,0036) | LO |                               | Always             |
| Visual Field Test Point Sequence                            | (0024,0089) | SQ |                               | Always             |

|                                                                    |             |    |            |                    |
|--------------------------------------------------------------------|-------------|----|------------|--------------------|
| > Visual Field Test Point XCoordinate                              | {0024,0090} | FL |            | Always             |
| > Visual Field Test Point YCoordinate                              | {0024,0091} | FL |            | Always             |
| > Stimulus Results                                                 | {0024,0093} | CS |            | Always             |
| > Sensitivity Value                                                | {0024,0094} | FL |            | Not always present |
| > Retest Stimulus Seen                                             | {0024,0095} | CS |            | Not always present |
| > Retest Sensitivity Value                                         | {0024,0096} | FL |            | Not always present |
| > Visual Field Test Point Normals Sequence                         | {0024,0097} | SQ |            | Not always present |
| > > AgeCorrectedSensitivityDeviation-Value                         | {0024,0092} | FL |            | Always             |
| > > AgeCorrectedSensitivityDeviation-ProbabilityValue              | {0024,0100} | FL |            | Always             |
| > > GeneralizedDefectCorrectedSensitivityDeviationFlag             | {0024,0102} | CS | "YES"/"NO" | Always             |
| > > GeneralizedDefectCorrectedSensitivityDeviationValue            | {0024,0103} | FL |            | Not always present |
| > > GeneralizedDefectCorrectedSensitivityDeviationProbabilityValue | {0024,0104} | FL |            | Not always present |

Table 7.7: Visual Field Static Perimetry Test Results

| Attribute Name                                             | Tag         | VR | Value                               | Presence of Value  |
|------------------------------------------------------------|-------------|----|-------------------------------------|--------------------|
| VisualFieldMeanSensitivity                                 | (0024,0070) | FL |                                     | Not always present |
| VisualFieldTestNormalsFlag                                 | (0024,0063) | CS |                                     | Always             |
| ResultsNormalsSequence                                     | (0024,0064) | SQ |                                     | Not always present |
| > GlobalDeviationFromNormal                                | (0024,0066) | FL |                                     | Always             |
| > GlobalDeviationProbabilityNormalsFlag                    | (0024,0059) | CS |                                     | Always             |
| > LocalizedDeviationFromNormal                             | (0024,0068) | FL |                                     | Always             |
| > LocalDeviationProbabilityNormalsFlag                     | (0024,0072) | CS |                                     | Always             |
| > Data Set Name                                            | (0024,0306) | LO | "OCULUS normative data"             | Not always present |
| > Data Set Version                                         | (0024,0307) | LO |                                     | Not always present |
| > Data Set Source                                          | (0024,0308) | LO | "OCU"                               | Not always present |
| ShortTermFluctuationCalculated                             | (0024,0074) | CS | "YES"/"NO"                          | Always             |
| ShortTermFluctuation                                       | (0024,0075) | FL |                                     | Not always present |
| ShortTermFluctuationProbabilityCalculated                  | (0024,0076) | CS | "NO"                                | Always             |
| CorrectedLocalizedDeviationFromNormalCalculated            | (0024,0078) | CS | "YES"/"NO"                          | Always             |
| CorrectedLocalizedDeviationFromNormal                      | (0024,0079) | FL |                                     | Not always present |
| CorrectedLocalizedDeviationFromNormalProbabilityCalculated | (0024,0080) | CS | "NO"                                | Always             |
| ResultsNormalsSequence                                     | (0024,0320) | SQ |                                     | Not always present |
| > DataObservationSequence                                  | (0024,0325) | SQ |                                     | Always             |
| >> ValueType                                               | (0040,A040) | CS | "CODE"                              | Always             |
| >> ConceptNameCodeSequence                                 | (0040,A043) | SQ |                                     | Always             |
| >>> Code Value                                             | (0008,0100) | SH | "1532180"                           | Always             |
| >>> Code Scheme Designator                                 | (0008,0102) | SH | "OCU"                               | Always             |
| >>> Code Meaning                                           | (0008,0104) | LO | "Glaucoma Asymmetric Test Analysis" | Always             |
| >> ConceptCodeSequence                                     | (0040,A168) | SQ | 7.24                                | Always             |

Table 7.8: Ophthalmic Patient Clinical Information and Test Lens Parameter

| Attribute Name                                        | Tag         | VR | Value | Presence of Value                        |
|-------------------------------------------------------|-------------|----|-------|------------------------------------------|
| OphthalmicPatientClinicalInformationLeft-EyeSequence  | (0024,0114) | SQ |       | Present if left or both eye measurement  |
| > RefractiveParametersUsedOnPatientSequence           | (0024,0112) | SQ |       | Always                                   |
| >> SphericalLensPower                                 | (0022,0007) | FL |       | Not always present                       |
| >> CylinderLensPower                                  | (0022,0008) | FL |       | Not always present                       |
| >> CylinderAxis                                       | (0022,0009) | FL |       | Not always present                       |
| > PupilSize                                           | (0046,0044) | FD |       | Always                                   |
| > PupilDilated                                        | (0022,000d) | CS |       | Always but empty                         |
| OphthalmicPatientClinicalInformation-RightEyeSequence | (0024,0115) | SQ |       | Present if right or both eye measurement |
| > RefractiveParametersUsedOnPatientSequence           | (0024,0112) | SQ |       | Always                                   |
| >> SphericalLensPower                                 | (0022,0007) | FL |       | Not always present                       |
| >> CylinderLensPower                                  | (0022,0008) | FL |       | Not always present                       |
| >> CylinderAxis                                       | (0022,0009) | FL |       | Not always present                       |
| > PupilSize                                           | (0046,0044) | FD |       | Always                                   |
| > PupilDilated                                        | (0022,000d) | CS |       | Always but empty                         |

#### 7.1.1.4 Common Modules

Table 7.9: Patient Module of Created SOP Instances

| Attribute Name       | Tag         | VR | Value         | Presence of Value      |
|----------------------|-------------|----|---------------|------------------------|
| Patient's Name       | (0010,0010) | PN | From worklist | Always                 |
| Patient's ID         | (0010,0020) | LO | From worklist | Always, might be empty |
| Patient's Birth Date | (0010,0030) | DA | From worklist | Always                 |
| Patient's Sex        | (0010,0040) | CS | From worklist | Always, might be empty |

Table 7.10: General Study Module of Created SOP Instances

| Attribute Name                     | Tag          | VR | Value                                                                         | Presence of Value      |
|------------------------------------|--------------|----|-------------------------------------------------------------------------------|------------------------|
| Study Date                         | (0008,0020)  | DA | From worklist if available                                                    | Always, might be empty |
| Study Time                         | (0008,0030)  | TM | From worklist if available                                                    | Always, might be empty |
| Accession Number                   | (0008,0050)  | SH | From worklist                                                                 | Always, might be empty |
| Referring Physician's Name         | (0008,0090)  | PN | From worklist                                                                 | Not always present     |
| Study Description                  | (0008,1030)  | LO | From attribute (0032,1060) in the worklist or entered by the user             | Always, might be empty |
| Procedure Code Sequence            | (0008,1032)  | SQ | From attribute (0032,1064) in the worklist or selected from a predefined list | Not always present     |
| > Code Value                       | (0008,0100)  | SH |                                                                               | Always                 |
| > Coding Scheme Designator         | (0008,0102)  | SH |                                                                               | Always                 |
| > Code Meaning                     | (0008,0104)  | LO |                                                                               | Always                 |
| Name of Physician(s) Reading Study | (0008,1060)  | PN | Selected from a predefined list                                               | Not always present     |
| Study Instance UID                 | (0020, 000D) | UI | Either from worklist entry or generated by the system                         | Always                 |
| Study ID                           | (0020, 0010) | SH | Generated by the system                                                       | Always, might be empty |

Table 7.11: General Equipment Module of Created SOP Instances

| Attribute Name                | Tag          | VR | Value                       | Presence of Value      |
|-------------------------------|--------------|----|-----------------------------|------------------------|
| Manufacturer                  | (0008,0070)  | LO | "OCULUS Optikgeraete GmbH"  | Always                 |
| Institution Name              | (0008,0080)  | LO | From the configuration file | Always, might be empty |
| Institution Address           | (0008,0081)  | ST | From the configuration file | Always, might be empty |
| Station Name                  | (0008,1010)  | SH | From the configuration file | Always, might be empty |
| Institutional Department Name | (0008,1040)  | LO | From the configuration file | Always, might be empty |
| Manufacturer's Model Name     | (0008,1090)  | LO | "Smartfield "               | Always                 |
| Device Serial Number          | (00018,1000) | LO | Serial number of the device | Always                 |
| Software Versions             | (0018,1020)  | LO | eg "1.29r03\ 1.29r03"       | Always                 |

Table 7.12: SC Equipment Module of Created SOP Instances

| Attribute Name                                     | Tag         | VR | Value                      | Presence of Value |
|----------------------------------------------------|-------------|----|----------------------------|-------------------|
| Modality                                           | (0008,0060) | CS | "OT"                       | Always            |
| Conversion Type                                    | (0008,0064) | CS | "WSD"                      | Always            |
| Secondary Capture Device ID                        | (0018,1010) | LO | Identifier, e.g. "56900"   | Always            |
| Secondary Capture Device Manufacturer              | (0018,1016) | LO | "OCULUS Optikgeraete GmbH" | Always            |
| Secondary Capture Device Manufacturer's Model Name | (0018,1018) | LO | "Smartfield "              | Always            |

Table 7.13: Private Data Module of Created SOP Instances

| Attribute Name                            | Tag         | VR | Value                      | Presence of Value |
|-------------------------------------------|-------------|----|----------------------------|-------------------|
| Private Creator                           | (0029,0010) | LO | "OCULUS Optikgeraete GmbH" | Always            |
| Original Measuring Data                   | (0029,1010) | OB | Depends on measured data   | Always            |
| Original Measuring Data Length            | (0029,1012) | UL | Depends on measured data   | Always            |
| Original Measuring Data (Raw Data)        | (0029,1020) | OB | Depends on measured data   | Not always        |
| Original Measuring Data (Raw Data) Length | (0029,1022) | UL | Depends on measured data   | Not always        |

### 7.1.1.5 Encapsulated PDF Modules

Table 7.14: Encapsulated Document Series Module of Created SOP Instances

| Attribute Name                           | Tag         | VR | Value                                                             | Presence of Value      |
|------------------------------------------|-------------|----|-------------------------------------------------------------------|------------------------|
| Modality                                 | (0008,0060) | CS | See Table 7.12                                                    |                        |
| Series Instance UID                      | (0020,00E)  | UI | Generated automatically                                           | Always                 |
| Series Number                            | (0020,0011) | IS | Sequential number increased for each series within a study        | Always                 |
| Performed Procedure Step Start Date      | (0040,0244) | DA | Date of examination                                               | Always                 |
| Performed Procedure Step Start Time      | (0040,0245) | TM | Time of examination                                               | Always                 |
| Performed Procedure Step ID              | (0040,0253) | SH | From attribute (0040,0009) in the worklist or entered by the user | Not always present     |
| Performed Procedure Step Description     | (0040,0254) | LO | Entered by the user                                               | Always, might be empty |
| Performed Protocol Code Sequence         | (0040,0260) | SQ | Selected from a predefined list                                   | Always                 |
| > Code Value                             | (0008,0100) | SH |                                                                   | Always                 |
| > Coding Scheme Designator               | (0008,0102) | SH |                                                                   | Always                 |
| > Code Meaning                           | (0008,0104) | LO |                                                                   | Always                 |
| Comments on the Performed Procedure Step | (0040,0280) | ST | Entered by the user                                               | Always, might be empty |

Table 7.15: Encapsulated Document Module of Created SOP Instances

| Attribute Name                     | Tag         | VR | Value                                                         | Presence of Value      |
|------------------------------------|-------------|----|---------------------------------------------------------------|------------------------|
| Instance Number                    | (0020,0013) | CS | Sequential number increased for each instance within a series | Always                 |
| Content Date                       | (0008,0023) | DA | Date of examination                                           | Always                 |
| Content Time                       | (0008,0033) | TM | Time of examination                                           | Always                 |
| Acquisition DateTime               | (0008,002A) | DT |                                                               | Always empty           |
| Burned In Annotation               | (0028,0301) | CS | "YES"                                                         | Always                 |
| Document Title                     | (0042,0010) | ST | Textual description of exported data                          | Always, might be empty |
| Concept Name Code Sequence         | (0040,A043) | SQ |                                                               | Always                 |
| MIME Type of Encapsulated Document | (0042,0012) | LO | "application/pdf"                                             | Always                 |
| Encapsulated Document              | (0042,0011) | OB | Depends on the content of the PDF document                    | Always                 |



Table 7.16: SOP Common Module of Created SOP Instances

| Attribute Name         | Tag         | VR | Value                         | Presence of Value |
|------------------------|-------------|----|-------------------------------|-------------------|
| Specific Character Set | {0008,0005} | CS | "ISO_IR 100"                  | Always            |
| Instance Creation Date | {0008,0012} | DA | Current date                  | Always            |
| Instance Creation Time | {0008,0013} | TM | Current time                  | Always            |
| SOP Class UID          | {0008,0016} | UI | 1.2.840.10008.5.1.4.1.1.104.1 | Always            |
| SOP Instance UID       | {0008,0018} | UI | Generated automatically       | Always            |

Draft

### 7.1.1.6 Secondary Capture Image Modules

Table 7.17: General Series Module of Created SOP Instances

| Attribute Name                           | Tag         | VR | Value                                                                         | Presence of Value      |
|------------------------------------------|-------------|----|-------------------------------------------------------------------------------|------------------------|
| Series Date                              | (0008,0021) | DA | Date the series started                                                       | Always                 |
| Series Time                              | (0008,0031) | TM | Time the series started                                                       | Always                 |
| Modality                                 | (0008,0060) | CS | See Table 7.12                                                                |                        |
| Performing Physician's Name              | (0008,1050) | PN | Selected from a predefined list                                               | Always, might be empty |
| Operator's Name                          | (0008,1070) | PN | Selected from a predefined list                                               | Always, might be empty |
| Body Part Examined                       | (0018,0015) | CS | "HEAD"                                                                        | Always                 |
| Protocol Name                            | (0018,1030) | LO | Generated automatically from performed examination type                       | Always, might be empty |
| Patient Position                         | (0018,5100) | CS |                                                                               | Never                  |
| Series Instance UID                      | (0020,000E) | UI | Generated automatically                                                       | Always                 |
| Series Number                            | (0020,0011) | IS | Sequential number increased for each examination performed                    | Always                 |
| Laterality                               | (0020,0060) | CS | "L" or "R"                                                                    | Always                 |
| Performed Procedure Step Start Date      | (0040,0244) | DA | Date of examination                                                           | Always                 |
| Performed Procedure Step Start Time      | (0040,0245) | TM | Time of examination                                                           | Always                 |
| Performed Procedure Step ID              | (0040,0253) | SH | From attribute (0040,0009) in the worklist or selected from a predefined list | Always, might be empty |
| Performed Procedure Step Description     | (0040,0254) | LO | Entered by the user                                                           | Always, might be empty |
| Performed Protocol Code Sequence         | (0040,0260) | SQ | Selected from a predefined list                                               | Always, might be empty |
| > Code Value                             | (0008,0100) | SH |                                                                               | Always                 |
| > Coding Scheme Designator               | (0008,0102) | SH |                                                                               | Always                 |
| > Code Meaning                           | (0008,0104) | LO |                                                                               | Always                 |
| Comments on the Performed Procedure Step | (0040,0280) | ST | Entered by the user                                                           | Always, might be empty |

Table 7.18: General Image Module of Created SOP Instances

| Attribute Name      | Tag         | VR | Value                                                         | Presence of Value |
|---------------------|-------------|----|---------------------------------------------------------------|-------------------|
| Image Type          | (0008,0008) | CS |                                                               | Never             |
| Content Date        | (0008,0023) | DA | Date of examination                                           | Always            |
| Content Time        | (0008,0033) | TM | Time of examination                                           | Always            |
| Instance Number     | (0020,0013) | IS | Sequential number increased for each instance within a series | Always            |
| Patient Orientation | (0020,0020) | CS |                                                               | Never             |

Table 7.19: General Pixel Module of Created SOP Instances

| Attribute Name             | Tag         | VR | Value                               | Presence of Value |
|----------------------------|-------------|----|-------------------------------------|-------------------|
| Samples Per Pixel          | (0028,0002) | US | 3                                   | Always            |
| Photometric Interpretation | (0028,0004) | CS | "RGB"                               | Always            |
| Planar Configuration       | (0028,0006) | US | 0                                   | Always            |
| Rows                       | (0028,0010) | US | Depends on the height of the image  | Always            |
| Columns                    | (0028,0011) | US | Depends on the width of the image   | Always            |
| Pixel Aspect Ratio         | (0028,0034) | IS |                                     | Never             |
| Bits Allocated             | (0028,0100) | US | 8                                   | Always            |
| Bits Stored                | (0028,0101) | US | 8                                   | Always            |
| High Bit                   | (0028,0102) | US | 7                                   | Always            |
| Pixel Representation       | (0028,0103) | US | 0                                   | Always            |
| Pixel Data                 | (7FE0,0010) | OB | Depends on the content of the image | Always            |

Table 7.20: SOP Common Module of Created SOP Instances

| Attribute Name         | Tag         | VR | Value                       | Presence of Value |
|------------------------|-------------|----|-----------------------------|-------------------|
| Specific Character Set | (0008,0005) | CS | "ISO_IR 100"                | Always            |
| SOP Class UID          | (0008,0016) | UI | "1.2.840.10008.5.1.4.1.1.7" | Always            |
| SOP Instance UID       | (0008,0018) | UI | Generated automatically     | Always            |

### 7.1.1.7 Ophthalmic Visual Field Static Perimetry Measurements Modules

Table 7.21: General Series Module of Created SOP Instances

| Attribute Name                           | Tag         | VR | Value                                                                         | Presence of Value      |
|------------------------------------------|-------------|----|-------------------------------------------------------------------------------|------------------------|
| Series Date                              | (0008,0021) | DA | Date the series started                                                       | Always                 |
| Series Time                              | (0008,0031) | TM | Time the series started                                                       | Always                 |
| Modality                                 | (0008,0060) | CS | OPV                                                                           | Always                 |
| Performing Physician's Name              | (0008,1050) | PN | Selected from a predefined list                                               | Always, might be empty |
| Operator's Name                          | (0008,1070) | PN | Selected from a predefined list                                               | Always, might be empty |
| Body Part Examined                       | (0018,0015) | CS | "HEAD"                                                                        | Always                 |
| Protocol Name                            | (0018,1030) | LO | Generated automatically from performed examination type                       | Always, might be empty |
| Patient Position                         | (0018,5100) | CS |                                                                               | Never                  |
| Series Instance UID                      | (0020,000E) | UI | Generated automatically                                                       | Always                 |
| Series Number                            | (0020,0011) | IS | Sequential number increased for each examination performed                    | Always                 |
| Laterality                               | (0020,0060) | CS | "L" or "R"                                                                    | Always                 |
| Performed Procedure Step Start Date      | (0040,0244) | DA | Date of examination                                                           | Always                 |
| Performed Procedure Step Start Time      | (0040,0245) | TM | Time of examination                                                           | Always                 |
| Performed Procedure Step ID              | (0040,0253) | SH | From attribute (0040,0009) in the worklist or selected from a predefined list | Always, might be empty |
| Performed Procedure Step Description     | (0040,0254) | LO | Entered by the user                                                           | Always, might be empty |
| Performed Protocol Code Sequence         | (0040,0260) | SQ | cf. Tables 7.22 and 7.23                                                      | Always, might be empty |
| > Code Value                             | (0008,0100) | SH |                                                                               | Always                 |
| > Coding Scheme Designator               | (0008,0102) | SH |                                                                               | Always                 |
| > Code Meaning                           | (0008,0104) | LO |                                                                               | Always                 |
| Comments on the Performed Procedure Step | (0040,0280) | ST | Entered by the user                                                           | Always, might be empty |

Table 7.22: Visual Field Static Perimetry Test Patterns

| Coding Scheme Designator<br>(0008,0102) | Code Value<br>(0008,0100) | Code Meaning (0008,0104)                       |
|-----------------------------------------|---------------------------|------------------------------------------------|
| DCM                                     | 111800                    | Visual Field 24-2 Test Pattern                 |
| DCM                                     | 111801                    | Visual Field 10-2 Test Pattern                 |
| DCM                                     | 111802                    | Visual Field 30-2 Test Pattern                 |
| DCM                                     | 111803                    | Visual Field 60-4 Test Pattern                 |
| DCM                                     | 111804                    | Visual Field Macula Test Pattern               |
| DCM                                     | 111805                    | Visual Field Central 40 Point Test Pattern     |
| DCM                                     | 111806                    | Visual Field Central 76 Point Test Pattern     |
| DCM                                     | 111807                    | Visual Field Peripheral 60 Point Test Pattern  |
| DCM                                     | 111808                    | Visual Field Full Field 81 Point Test Pattern  |
| DCM                                     | 111809                    | Visual Field Full Field 120 Point Test Pattern |
| DCM                                     | 111810                    | Visual Field G Test Pattern                    |
| DCM                                     | 111811                    | Visual Field M Test Pattern                    |
| DCM                                     | 111812                    | Visual Field O7 Test Pattern                   |
| DCM                                     | 111813                    | Visual Field LVC Test Pattern                  |
| DCM                                     | 111814                    | Visual Field Central Test Pattern              |
| OCU                                     | 1532111                   | Visual Field Esterman Test Pattern             |
| OCU                                     | 1532112                   | Visual Field FeV-70 Test Pattern               |
| OCU                                     | 1532113                   | Visual Field 30x24 Test Pattern                |
| OCU                                     | 1532131                   | Visual Field Area1 Test Pattern                |
| OCU                                     | 1532132                   | Visual Field Area2 Test Pattern                |
| OCU                                     | 1532133                   | Visual Field Area3 Test Pattern                |
| OCU                                     | 1532134                   | Visual Field Area4 Test Pattern                |
| OCU                                     | 1532135                   | Visual Field Area5 Test Pattern                |
| OCU                                     | 1532136                   | Visual Field Area6 Test Pattern                |

Table 7.23: Visual Field Static Perimetry Test Strategies

| Coding Scheme Designator<br>(0008,0102) | Code Value<br>(0008,0100) | Code Meaning (0008,0104)                    |
|-----------------------------------------|---------------------------|---------------------------------------------|
| DCM                                     | 111818                    | Visual Field Full Threshold Test Strategy   |
| DCM                                     | 111822                    | Visual Field Two-Zone Test Strategy         |
| DCM                                     | 111823                    | Visual Field Three-Zone Test Strategy       |
| DCM                                     | 111824                    | Visual Field Quantify-Defects Test Strategy |
| DCM                                     | 111835                    | Visual Field Fast Threshold Test Strategy   |
| DCM                                     | 111836                    | Visual Field CLIP Test Strategy             |
| DCM                                     | 111837                    | Visual Field Class Test Strategy            |
| OCU                                     | 1532140                   | Visual Field SPARK Training Test Strategy   |
| OCU                                     | 1532141                   | Visual Field SPARK Quick Test Strategy      |
| OCU                                     | 1532142                   | Visual Field SPARK Precision Test Strategy  |
| OCU                                     | 1532143                   | Visual Field SPARKN Training Test Strategy  |
| OCU                                     | 1532144                   | Visual Field SPARKN Quick Test Strategy     |
| OCU                                     | 1532145                   | Visual Field SPARKN Precision Test Strategy |

Table 7.24: Visual Field Static Perimetry Test Analysis Results

| Coding Scheme Designator<br>(0008,0102) | Code Value<br>(0008,0100) | Code Meaning (0008,0104)                        |
|-----------------------------------------|---------------------------|-------------------------------------------------|
| DCM                                     | 111847                    | Outside normal limits                           |
| DCM                                     | 111848                    | Borderline                                      |
| DCM                                     | 111849                    | Abnormally high sensitivity                     |
| DCM                                     | 111850                    | General reduction in sensitivity                |
| DCM                                     | 111851                    | Borderline and general reduction in sensitivity |
| SRT                                     | M-00101                   | Within normal limit                             |

Table 7.25: Visual Field Static Perimetry Fixation Strategy

| Coding Scheme Designator<br>(0008,0102) | Code Value<br>(0008,0100) | Code Meaning (0008,0104) |
|-----------------------------------------|---------------------------|--------------------------|
| DCM                                     | 111844                    | Blind Spot Monitoring    |
| DCM                                     | 111845                    | Macular Fixation Testing |
| DCM                                     | 111846                    | Observation by Examiner  |

Table 7.26: Visual Field Illumination Color

| Coding Scheme Designator<br>(0008,0102) | Code Value<br>(0008,0100) | Code Meaning (0008,0104) |
|-----------------------------------------|---------------------------|--------------------------|
| SRT                                     | G-A11D                    | Yellow                   |
| SRT                                     | G-A12B                    | White                    |
| SRT                                     | G-A11A                    | Red                      |
| SRT                                     | G-A12F                    | Blue                     |
| SRT                                     | G-A11E                    | Green                    |

Draft

### 7.1.2 Usage of Attributes from Received IODs

OCULUS Smartfield does not receive any IODs.

### 7.1.3 Attribute Mapping

OCULUS Smartfield does not use any attribute mapping.

### 7.1.4 Coerced / Modified Fields

OCULUS Smartfield does not modify any fields.

## 7.2 Data Dictionary of Private Attributes

The private attributes added to created SOP instances are listed in Table 7.13. OCULUS Smartfield reserves a block of private attributes in the range of (0029,1000) to (0029,10FF).

## 7.3 Coded Terminology and Templates

OCULUS Smartfield does not use or support any coded terminology or templates.

## 7.4 Grayscale Image Consistency

OCULUS Smartfield does not make use of the DICOM Grayscale Standard Display Function.

## 7.5 Standard Extended / Specialized / Private SOP Classes

OCULUS Smartfield uses a standard extended SOP class for the Encapsulated PDF and Secondary Capture Image IOD by adding a private data module. See 7.1, page 22 for details. The system does not support any specialized or private SOP classes.

## 7.6 Private Transfer Syntaxes

OCULUS Smartfield does not use or support any private transfer syntaxes.



Draft

**WWW.OCULUS.DE**

**OCULUS Optikgeräte GmbH**

Postfach • 35549 Wetzlar • GERMANY  
Tel. +49 641 2005-0 • Fax +49 641 2005-255  
Email: sales@oculus.de • www.oculus.de

**OCULUS USA:**

OCULUS Inc.  
17721 59th Avenue NE • Arlington • WA 98223  
Tel. +1 425 670 9977 • Fax +1 425 670 0742  
Email: sales@oculususa.com • www.oculususa.com